

ian_Journal_of_Chemistry,_Vol.1
8,_No.2,_2018_,_Pages_279-
285.pdf
by

Submission date: 27-Jun-2023 05:59PM (UTC+0700)

Submission ID: 2123424773

File name: ian_Journal_of_Chemistry,_Vol.18,_No.2,_2018_,_Pages_279-285.pdf (246.09K)

Word count: 4116

Character count: 20219

The Utilization of Nitrogen Gas as a Carrier Gas in the Determination of Hg Ions Using Cold Vapor-Atomic Absorption Spectrophotometer (CV-AAS)

Aman Sentosa Panggabean*, Subur P. Pasaribu, and Farida Kristiana

Department of Chemistry, Faculty of Mathematics and Natural Sciences, Mulawarman University,
Jl. Barong Tongkok No.4 Kampus Gn. Kelua Samarinda 75123, Indonesia

Received March 17, 2017; Accepted December 14, 2017

ABSTRACT

The research about utilization of nitrogen gas as a carrier gas in the determination of Hg ions by using Cold Vapor-Atomic Absorption Spectrophotometer (CV-AAS) method has been conducted. To optimize the measurement results, several parameters that affect hydride generator have been studied. Some specified important parameters are SnCl_2 concentration as reductant, acid concentration, and the analytical performance such as repeatability and reproducibility (% RSD), linearity (r), limits of detection (LOD) and limits of quantitation (LOQ), and accuracy have been studied. The results of the research showed that the nitrogen gas can be used instead of argon gas as a carrier gas. It was shown by the repeatability values as % RSD < 2/3 CV Horwitz values, the LOD was 0.0338 $\mu\text{g/L}$ and the LOQ was 0.0838 $\mu\text{g/L}$. The accuracy of this method was well shown by a recovery percentage of 102.24%. Based on the result of this research, the nitrogen gas can be used as a carrier gas for the determination of Hg ions by using the CV-AAS method with valid results.

Keywords: Hg; carrier gas; nitrogen; CV-AAS

ABSTRAK

Penelitian tentang pemanfaatan gas nitrogen sebagai gas pembawa dalam penentuan ion Hg dengan menggunakan metode Bejana Uap Dingin-Spektrofotometer Serapan atom (CV-AAS) telah dilakukan. Untuk mengoptimalkan hasil pengukuran, beberapa parameter yang berpengaruh dalam pembangkit hidrida telah dipelajari. Beberapa parameter penting yang ditentukan adalah konsentrasi SnCl_2 sebagai reduktor, konsentrasi asam, dan kinerja analitik yang meliputi penentuan kedapatulangan dan kebolehlulangan (% RSD), linearitas (r), limit deteksi (LOD) dan limit kuantitasi (LOQ), dan akurasi. Hasil penelitian menunjukkan gas nitrogen dapat digunakan sebagai pengganti gas argon sebagai gas pembawa, ditunjukkan dari nilai kedapatulangan sebagai % RSD < 2/3 nilai KV Horwitz, LOD sebesar 0,0338 mg/L dan LOQ adalah 0,0838 mg/L. Ketepatan metode ini sangat baik yang ditunjukkan dengan persentase perolehan kembali 102,24%. Berdasarkan hasil penelitian ini, gas nitrogen dapat digunakan sebagai gas pembawa untuk penentuan ion Hg dengan menggunakan metode CV-AAS dengan hasil yang valid.

Kata Kunci: Hg; gas pembawa; nitrogen; CV-AAS

INTRODUCTION

Mercury (Hg) is a toxic element that can be found in different chemical forms at trace levels and it is very reactive in the environment. Hg is obtained in nature and distributed throughout the environment in the inorganic and organic forms (for example methylmercury) of the element [1]. Hg can be found mostly in the atmosphere, water, soil, sediments, plants, pharmaceuticals, industrial waste, mining materials and animals. The main indirect source of human contamination of Hg is through the consumption of fish products [2-3]. Undoubtedly, the natural occurrence of Hg can be found in crude oil and its presence in fuels [4]. Thus, the need for the development and validation of analytical methodologies

for the determination and control of this toxic metal in samples is evident.

Several methods for the determination of Hg have been reported in the literature, such as potentiometric titration using iodide ion-selective electrodes and anodic stripping voltammetry [5], and some spectrometric techniques, such as cold vapor atomic absorption spectrometry (CV-AAS) [6-7], electrothermal atomic absorption spectrometry (ETAAS) [8], atomic fluorescence spectrometry (AFS) [9], inductively coupled plasma mass spectrometry (ICP-MS) [10] and electrothermal vaporization coupled to ICP-MS (ETV-ICP-MS) [11] were already employed for total Hg determination in the samples.

* Corresponding author.
Email address : amanspanggabean@yahoo.com

The hydride generation of cold vapor of mercury coupled to spectrometric techniques is usually employed. Cold vapor-atomic absorption spectrophotometry (CV-AAS) is the most widely used technique for the determination of Hg because of its main advantages including the possibility of separation of the analyte from its matrix with the inherent minimization of matrix interferences and the possibility of analyte pre-concentration, leading to increased measurement and the results obtained even at the low levels of concentration [12-14].

The fundamental principle of determination of Hg in the samples is the formation of metal-hydride by reductant SnCl_2 or NaBH_4 in gas-liquid separator placed in hydride generator system and produce Hg-hydride. The formed Hg-hydride will then go to a quartz cell placed on either the AAS is driven by a carrier gas of argon to read absorption value [15-16]. The atomization occurred in the quartz cell is caused by the light from the hollow cathode that passes it through. The advantage of this method is that the detection limit and sensitiveness can be enhanced because the metal has been separated from each other in their hydride forms [17].

Analysis of Hg in Laboratory of Environment Control, P.T. Badak NGL-East Kalimantan has been using the CV-AAS method with argon as a carrier gas. Argon gas is an inert gas that has high enough purity, but quite expensive and takes times because it must be purchased from a partner. Frequent reliance on the availability of argon gas, if there is no supply of argon gas, analysis cannot be done immediately. Based on, raised a new idea for replacing the carrier gas with cheaper gas and does not depend on the procurement process. In this case, nitrogen gas is already available in the laboratory because it is one of the products of the utility unit in P.T. Badak NGL refinery and used for the LNG/LPG process usage.

In this research, it has been developed a method using nitrogen gas as a carrier gas. Nitrogen gas is obtained from the refinery processes, namely P.T. Badak NGL #29th plant utility. Therefore, a study to determine the effect of utilization of nitrogen as a carrier gas in the determination of Hg using CV-AAS method needs to be conducted as well as to determine the validity of the developed method for appropriate used in the analysis.

EXPERIMENTAL SECTION

Materials

A solution of CRM Mercury in Water (Low Level PT-2) RTC Number: CS-1310 LOT: 017 937 with a concentration value of $2.48 \pm 0.04 \mu\text{g/L}$ and Acceptance Limits of 2.15–2.81 $\mu\text{g/L}$, Hg stock solution 1000 mg/L,

SnCl_2 (concentration series of 6-14% w/v), HNO_3 , HCl, nitrogen gas; all of the reagents were of analytical-reagent grade (E. Merck), and double-distilled water.

Instrumentation

Spectrophotometer AA 240-FS variant complete with Vapor Generator Accessory VGA-76. Other equipment such as analytical balance, the volume pipettes, stopwatch, oven, hot plate with stirrer, was used for all measurements.

Procedure

Quality determination of nitrogen gas

Nitrogen gas was stored in pressurized gas cylinder mounted regulator and equipped with a filter and moisture trap, then analyzed using gas chromatography instrument.

Preparation of blank solution

The prepared blank solution is distilled water containing 5% (v/v) HCl. Double-distilled water 300 mL was entered to 500 mL volumetric flask, add 65 mL 37% HCl and matched in volume with double-distilled water, whipped until homogeneous.

Preparation of Hg stock solution 1000 mg/L

Amount of 0.1354 g HgCl_2 was diluted with 70 mL double-distilled water, added with 1 mL HNO_3 and it was diluted to 100 mL with double distilled water.

Preparation of CRM mercury standard solutions in Water (Low-Level PT-2) RTC Number: CS-1310 LOT: 017 937

CRM standard solution of 10 mL was piped and diluted in 1000 mL double-distilled water and 1 mL HNO_3 . This solution has a concentration value of $2.48 \pm 0.04 \mu\text{g/L}$ and acceptance limits 2.15–2.81 $\mu\text{g/L}$ accordance with the value stated in the Certificate of Analysis.

Preparation of series Hg CRM standard solution in Water (Low-Level PT-2) RTC Number: CS-1310 LOT: 017 937

A certain volume of Hg CRM standard solution was diluted in double-distilled water with adding concentrated HNO_3 to obtain series of concentration (1.24, 1.86, 2.48, and 3.72 $\mu\text{g/L}$).

Determination of reading delay time

Read delay time was measured for 0–60 sec after standard solution added into mixing system (in the reaction coil). The maximum absorbance was

subsequently used to the analytical performance test methods of Hg determination.

Determination of SnCl₂ concentration as reductant

Series concentrations of SnCl₂ (5 to 12% (w/v)) was prepared and measured their absorption. Maximum absorbance then subsequently used to the analytical performance test methods of Hg determination.

Determination of HCl concentration

Concentration of HCl (2 to 7% (v/v)) solutions were measured their absorption. Maximum absorbance then subsequently used to the analytical performance test methods of Hg determination.

Calibration curves and regression line equation

Hg standard with series concentrations of 0.0 - 12.0 µg/L and Hg CRM concentrations of 0:00; 1:24; 1.86; 2:48; 3.72 µg/L were measured by CV-AAS to obtain absorbance values.

Repeatability test

Repeatability test is performed by 10 times measuring standard solutions of Hg 6.0 and 9.0 µg/L and Hg CRM standard solutions in Water (Low-Level PT-2) RTC Number: CS-1310 LOT: 017 937.

Accuracy test

Accuracy test was done in two ways: first, by measuring 10 times Hg standard solutions of 6.0 and 9.0 µg/L then calculated % recovery. The second way is by measuring 10 times Hg Certified Reference Material (CRM) solution by CV-AAS, then calculating the average and compared with specified Certified Value in the Certificate of Analysis Standards.

Limit of detection and quantization

Detection and quantitation limits were determined by measuring double-distilled water reference solution with spike until Hg levels up to 0.1 mg/L, and 5% (v/v) HCl and analyzed 10 times by CV-AAS.

RESULT AND DISCUSSION

Nitrogen Gas Quality Analysis

Nitrogen gas obtained from PT Badak NGL #29th plant and analyzed first by using HP 6890 Agilent gas chromatography before it is used as a carrier gas in Hg analysis by the CV-AAS method. Gas chromatography (GC) is a widely used method for determining the composition of complex mixtures of chemical species in the gaseous phase. In the course of a GC analysis, the components of the mixture are separated on the basis of differences in the rate at which they are carried through a




Fig 1. Reading delay time variation chart

stationary phase by a mobile gas phase [18]. The results show that nitrogen gas has very high purity up to 99.9495%.

Read Delay Time

Read delay time is a required time for Hg vapor flowing through VGA pipe tube of reaction coil to absorption cells until absorbance signal reading is done. The reaction coil is a mixing place of the sample, H⁺ and SnCl₂ reductant, Hg²⁺ ions are reduced to Hg⁰ and free atoms Hg⁰ gaseous is going to absorption cell for absorbance measurement [16]. Time delay greatly affects the low and high of absorbance. Determination of the delay time by measurement of absorbance from Hg standards 6.00; 9.00 µg/L and Hg CRM 2.48 ± 0.04 µg/L with a time delay variation 5–50 sec as shown in Fig. 1. The optimum absorbance of reading time delay was in 20 sec.

Influence of SnCl₂ Concentration as Reductant

The generation of Hg cold vapor depends on the addition of a suitable amount of reducing agent, since mercury is found in solution in the form of Hg(II) ions and needs to be reduced to gaseous elemental Hg [6]. SnCl₂ was employed for this purpose. Determination of the influence of SnCl₂ concentration in reducing Hg²⁺ ion into Hg⁰ was performed by optimization test with a varying SnCl₂ concentration of 5–12% (w/v) with H⁺ and Hg standard concentration made permanent. The results of research shown in Fig. 2. The optimal concentration of SnCl₂ gives a maximum absorbance of




Fig 2. Influence of SnCl₂ concentration




Fig 3. Influence of HCl concentration





Fig 4. Calibration curve of (a) Hg standard and (b) Hg CRM

Hg CRM $2.48 \pm 0.04 \mu\text{g/L}$ was 10% (w/v). Therefore, a 10% (w/v) SnCl₂ solution was employed in all further experiments.

Influence of Acid Concentration

The vapor generation of Hg also depends on the acidity of the medium since the decomposition of SnCl₂ into hydrostanno intermediates, which are responsible for the reduction of Hg(II) ions, is accelerated in acidic medium [19]. Hence, to keep the medium acid sufficient for promoting suitable Hg vapor evolution, HCl solution was added to the sample (or standard solution) before addition of SnCl₂ solution. In the research stage, HCl as acids prepared in several concentrations 2–7% (v/v), while standard concentrations of Hg and SnCl₂ reductant made permanent. The results obtained in this experiment (Fig. 3) showed that the highest response

could be achieved when HCl concentration was 5% (v/v) with produced to highest Hg CRM $2.48 \pm 0.04 \mu\text{g/L}$. When the HCl concentration was higher than this value, a decrease of Hg signals was verified. In view of these results, a 5% (v/v) HCl solution was selected for the method.

Determination of Calibration Curves

Determination of calibration curve and the regression line is done by measuring the Hg standard and Hg CRM in various concentrations. Measurement results are shown in Fig. 4. The concentrations of Hg standards are in the range 1.0–12.0 $\mu\text{g/L}$. The linear regression equation: $y = 0.0154x + 0.0013$ with coefficient of correlation (R^2) = 0.9997 (Fig. 4a). Hg CRM calibration data is shown in Fig. 4b can be seen that the area of the linear range of concentration CRM

Table 1. Measurement of Hg standard and Hg CRM for repeatability test

Measurement	Hg Standard 6.0 $\mu\text{g/L}$	Hg Standard 9.0 $\mu\text{g/L}$	Hg CRM 2.48 \pm 0.04 $\mu\text{g/L}$
2	6.162	9.265	2.5236
3	6.172	9.283	2.4919
4	6.114	9.258	2.5101
5	6.133	9.268	2.5248
6	6.132	9.290	2.5098
7	6.125	9.274	2.4917
8	6.113	9.272	2.5207
9	6.147	9.279	2.4910
10	6.138	9.248	2.5079
Mean (X)	6.134	9.271	2.5083
SD	0.022	0.012	0.0138
% RSD	0.351	0.131	0.5495
CV Horwitz	34.443	32.367	39.4050
2/3 CV Horwitz	22.962	21.578	26.2700

Table 2. Measurement of Hg standard and CRM solution

Measurement	Hg 6.0 $\mu\text{g/L}$		Hg 9.0 $\mu\text{g/L}$		Hg CRM 2.48 \pm 0.04 $\mu\text{g/L}$	
	Measured	% Recovery	Measured	% Recovery	Measured	% Recovery
1	6.105	101.75	9.270	103.00	2.5115	101.27
2	6.162	102.70	9.265	102.94	2.5236	101.76
3	6.172	102.87	9.283	103.14	2.4919	100.48
4	6.114	101.90	9.258	102.82	2.5101	101.21
5	6.133	102.22	9.268	102.98	2.5248	101.81
6	6.132	102.20	9.290	103.22	2.5098	101.20
7	6.125	102.08	9.274	103.04	2.4917	100.47
8	6.113	101.88	9.272	103.02	2.5207	101.64
9	6.147	102.45	9.279	103.10	2.4910	100.44
10	6.138	102.30	9.248	102.76	2.5196	101.60
Mean \pm SD	6.134 \pm 0.022	102.24 \pm 0.36	9.271 \pm 0.012	103.01 \pm 0.14	2.5095 \pm 0.013	101.19 \pm 0.55

Hg is 1.24–3.72 $\mu\text{g/L}$, the obtained the linear regression equation are $y = 0.0135x - 0.0001$ with a coefficient of correlation (R^2) = 0.9986. The measurement results of Hg standard and Hg CRM very good for Hg analysis, shown by the coefficient of correlation (R^2) are > 0.995 [20-21].

Repeatability Measurement

Repeatability is determined by measuring the concentration of Hg standard solution 6.0 and 9.0 $\mu\text{g/L}$ and Hg CRM solution 2.48 \pm 0.04 $\mu\text{g/L}$ repeatedly with optimum conditions. Table 1 showed, % RSD of Hg standard solution 6.0 $\mu\text{g/L}$ was 0.351% $<$ 22.962, % RSD of Hg standard 9.0 $\mu\text{g/L}$ was 0.131% $<$ 21.578 and % RSD of Hg CRM solution 2.48 \pm 0.04 $\mu\text{g/L}$ was 0.5495% $<$ 26.2700. All of % RSD of each Hg standard $<$ 2/3 of CV Horwitz. This result indicated the good repeatability and acceptable measurement.

Determination of Accuracy

Accuracy is the closeness of analysis data with actual values. Accuracy value is determined in two ways: the first way is to measure the concentration of Hg

standard solution 6.0 and 9.0 $\mu\text{g/L}$, and then % recovery value was calculated. Table 2 showed % recovery measurement of Hg standard 9.0 $\mu\text{g/L}$ was 103.01 \pm 0.14 and % recovery measurement of standard 6.0 $\mu\text{g/L}$ was 102.24 \pm 0.36. Each Hg standard result was in the range of required recommendation % recovery is 95–105% [21], and the accuracy was accepted.

The second way is to measure standard of Hg CRM 2.48 \pm 0.04 $\mu\text{g/L}$ repeatedly, then by comparing the mean of measurements with CRM values and acceptance limits stated in the certificate of CRM, which was shown in Table 2. The Mean Value of CRM was 2.5095 \pm 0.0130 $\mu\text{g/L}$ with % recovery was 101.19 \pm 0.55 and appropriate in the value of CRM certification limits 2.48 \pm 0.0383 $\mu\text{g/L}$ and accordance within the concentrations range of CRM acceptance limits were 2.4417 to 2.5183 $\mu\text{g/L}$, and the accuracy was accepted.

Determination of Limit of Detection and Quantization

Limit of detection (LOD) is determined by measuring the smallest absorbance value which still can be defined and distinguished from signals given by

Table 3. Measurement of blank solution

Measurement [Hg] (0.10 µg/L)	Absorbance	Hg Concentration (µg/L)
1	0.0015	0.0130
2	0.0013	0.0000
3	0.0015	0.0130
4	0.0015	0.0130
5	0.0015	0.0130
6	0.0015	0.0130
7	0.0014	0.0065
8	0.0014	0.0065
9	0.0016	0.0195
10	0.0017	0.0260
Mean (X)		0.0123
Standard Deviation (SD)		0.0071
LoD = X + (3xSD)		0.0338
LoQ = X + (10xSD)		0.0838

blank with some measurements [22]. Table 3 showed that the limit of detection was 0.0338 µg/L and the limit of quantitation was 0.0838 µg/L. This result showed the lowest concentration of sample that can still be detected by instruments and the analysis methods was 0.0338 µg/L while the lowest concentration of analyte that can be determined with acceptable accuracy was 0.0838 µg/L.

CONCLUSION

The results of the research showed that the nitrogen gas can be used as a carrier gas replacing argon gas in the determination of Hg ions, which are shown the repeatability values as % RSD < 2/3 CV Horwitz values, the LOD was 0.0338 µg/L and the LOQ was 0.0838 µg/L. The accuracy of this method is good shown by a recovery percentage was 102.24%, the CRM readings are 2.5095 µg/L and accordance within the CRM certificate of the range 2.48 ± 0.04 µg/L. Based on this research, the nitrogen gas can be used as a carrier gas for the determination of Hg ions in the samples by using CV-AAS with valid results.

ACKNOWLEDGEMENT

The authors would like to thank the Chairman and staff of P.T. Badak NGL Bontang, East Kalimantan for the laboratory facilities and the samples provided in the implementation of this research.

REFERENCES

- [1] Li, P., Feng, X., and Qiu, G., 2010, Methylmercury exposure and health effects from rice and fish consumption: A review, *Int. J. Environ. Res. Public Health*, 7 (6), 2666–2691.
- [2] Chen, Y., Dong, X., Dai, Y., Hu, Q., and Yu, H., 2008, Determination of trace mercury in Chinese

herbal medicine by cold vapour generation atomic fluorescence spectrometry, *Asian J. Chem.*, 20 (6), 4639–4646.

- [3] Tessier, A., and Turner, D.R., 1995, *Metal Speciation and Bioavailability in Aquatic Systems*, IUPAC Series on Analytical and Physical Chemistry of Environmental System, Vol. 3, John Wiley and Sons Ltd, Chichester, 679.
- [4] Shpirt, M.Ya., and Punanova, S.A., 2011, Accumulation of mercury in petroleum, coal, and their conversion products, *Solid Fuel Chem.*, 45, 330–336.
- [5] do Nascimento, F.H., and Masini, J.C., 2012, Complexation of Hg(II) by humic acid studied by square wave stripping voltammetry at screen-printed gold electrodes, *Talanta*, 100, 57–63.
- [6] Vicentino, P.O., Brum, D.M., and Cassella, R.J., 2015, Development of a method for total Hg determination in oil samples by cold vapor atomic absorption spectrometry after its extraction induced by emulsion breaking, *Talanta*, 132, 733–738.
- [7] Oreste, E.Q., de Jesus, A., de Oliveira, R.M., da Silva, M.M., Vieira, M.A., and Ribeiro, A.S., 2013, New design of cold finger for sample preparation in open system: Determination of Hg in biological samples by CV-AAS, *Microchem. J.*, 109, 5–9.
- [8] dos Santos, W.N.L., Dias, F.S., Reboucas, M.V., Pereira, M.G., Lemos, V.A., and Teixeira, L.S.G., 2006, Mercury determination in petroleum products by electrothermal atomic absorption spectrometry after in situ preconcentration using multiple injections, *J. Anal. At. Spectrom.*, 21 (11), 1327–1330.
- [9] Uddin, R., Al-Fahad, M.A., Al-Rashwan, A.K., and Al-Qarni, M.A., 2013, A simple extraction procedure for determination of total mercury in crude oil, *Environ. Monit. Assess.*, 185 (5), 3681–3685.

- [10] Chen, F.Y., and Jiang, S.J., 2009, Determination of Hg and Pb in fuels by inductively coupled plasma mass spectrometry using flow injection chemical vapor generation, *Anal. Sci.*, 25 (12), 1471–1476.
- [11] SaintPierre, T.D., Rocha, R.C.C., and Duyck, C.B., 2013, Determination of Hg in water associate to crude oil production by electrothermal vaporization inductively coupled plasma mass spectrometry, *Microchem. J.*, 109, 41–45.
- [12] Caiminagua, A., Fernández, L., Romero, H., Lapo, B., and Alvarado, J., 2015, Electrochemical generation of arsenic volatile species using a gold/mercury amalgam cathode. Determination of arsenic by atomic absorption spectrometry, *Anal. Chem. Res.*, 3, 82–88.
- [13] Kumar, A.R., and Riyazuddin, P., 2005, Mechanism of volatile hydride formation and their atomization in hydride generation atomic absorption spectrometry, *Anal. Sci.*, 21 (12), 1401–1410.
- [14] Silva, L.O.B., Leao, D.J., dos Santos, D.C., Matos, G.D., de Andrade, J.B., and Ferreira, S.L.C., 2014, Determination of copper in airborne particulate matter using slurry sampling and chemical vapor generation atomic absorption spectrometry, *Talanta*, 127, 140–145.
- [15] Thompson, K.C., and Thomerson, D.R., 1974, Atomic absorption studies on the determination of antimony, arsenic, bismuth, germanium, lead, selenium, tellurium and tin by utilising the generation of covalent hydrides, *Analyst*, 99 (1182), 595–601.
- [16] Nakahara, T., 2005, Development of gas-phase sample-introduction techniques for analytical atomic spectrometry, *Anal. Sci.*, 21 (5), 477–484.
- [17] Panggabean, A.S., Pasaribu, S.P., Amran, M.B., and Buchari, B., 2013, Gas-liquid separator integrated to HG-GFAAS method for determination of tin at trace levels in the water samples, *Eurasian J. Anal. Chem.*, 8 (1), 17–27.
- [18] Dantas, H.V., Barbosa, M.F., Moreira, P.N.T., Galvão, R.K.H., and Araújo, M.C.U., 2015, An automatic system for accurate preparation of gas mixtures, *Microchem. J.*, 119, 123–127.
- [19] D'Ulivo, A., 2004, Chemical vapor generation by tetrahydroborate(III) and other borane complexes in aqueous media: A critical discussion of fundamental processes and mechanisms involved in reagent decomposition and hydride formation, *Spectrochim. Acta, Part B*, 59 (6), 793–825.
- [20] Ritschdorff, E.T., Fitzgerald, N., McLaughlin, R.G.J., and Brindle, I.D., 2005, The use of a modified Multimode Sample Introduction System for the simple and rapid determination of cadmium by chemical vapour generation atomic absorption spectrometry, *Spectrochim. Acta, Part B*, 60 (1), 139–143.
- [21] Miller, J.C., and Miller, J.N., 1991, *Statistika Untuk Kimia Analitik*, 2nd ed., ITB Press, Bandung.
- [22] Panggabean, A.S., Pasaribu, S.P., Bohari, B., and Nurhasanah, H., 2014, Preconcentration of Chromium(VI) at trace levels using acid alumina resin with column method, *Indones. J. Chem.*, 14 (1), 51–56.

ORIGINALITY REPORT

17%

SIMILARITY INDEX

8%

INTERNET SOURCES

14%

PUBLICATIONS

3%

STUDENT PAPERS

MATCH ALL SOURCES (ONLY SELECTED SOURCE PRINTED)

3%

★ Mansoor Khan, Mustafa Soylak. "Deep Eutectic Solvent Based Liquid-Liquid Microextraction of Mercury in Water, Hair and Fish with Spectrophotometric Determination: A Green Protocol", Analytical Letters, 2022

Publication

Exclude quotes Off

Exclude matches Off

Exclude bibliography On